

MEDICAL SURGICAL MASK



MADE IN SINGAPORE



ASTM
F2100-11
LEVEL 2



- Registered with Health Science Authority of Singapore
- Size: 175mm x 95mm
- Test reports are done by TUV SUD PSB Singapore
- ASTM F2100-11 Level 2 Standard
- Bacterial Filtration Efficiency (BFE) > 99.96%
- Sub-Micron Particulate Filtration Efficiency (PFE) at 0.1 micron > 99.92%
- Synthetic Blood Penetration Test: Passed @ 120mmHg
- Breathability Test (Differential Pressure) : Passed Level 1 Barrier of less than 5mm H₂O/cm²
- Flammability Test: Passed Class 1 requirement
- Microbial Cleanliness: below 6 cfu/g
- Manufactured under ISO 13485:2016 standard
- Expiry Date: 2 years from date of manufacture
- Made of 3-ply of non-woven materials
- Comfortable & Perfect Fitting
- Ear Loop Design & Nose Bar Adaptable
- High Filtration Capacity
- Hypoallergenic



For inquiries, call **6490 5000** or reach us at customerservice@limkimhai.com.sg